REQUEST FOR QUOTATION

To,
Whomsoever it may concern

Date: 18 November 2026

Subject: Request for Quotation for Supply, Transportation, Installation and Commissioning of Medical Equipment and Hospital Furniture for Hospitals in Central India

Dear Sir/Madam,

BharatCares invites you to submit your quotation for the supply, transportation, installation, commissioning and demonstration of medical equipment and hospital furniture for the hospitals listed below under the CSR project. The hospital-wise equipment requirements, preferred brands, specifications, warranty requirements and quantities are detailed in Annexure-I.

Terms and Conditions

1. All equipment must be new, unused, and supplied with all standard accessories and required software, wherever applicable.
2. Vendors must quote the make, model, unit price, applicable GST and total price for each item.
3. The quotation must include transportation, delivery, installation and commissioning charges, wherever applicable. Any additional charges must be clearly itemised.
4. The quotation must remain valid for a minimum of 60 days from the date of submission.
5. The quotation must be duly signed and stamped by the authorised representative of the vendor.
6. The selection will be based on technical compliance, price, delivery timeline, warranty and after-sales service.
7. The purchaser reserves the right to accept or reject any or all quotations in accordance with its procurement policy.
8. Any deviation from the specifications, preferred brand or warranty requirement must be clearly stated against the relevant item.

Submission of Quotation

Interested vendors are requested to submit their signed and stamped quotations along with relevant supporting documents (PAN card, Aadhaar card, GST certificate, relevant work samples/experience references and cancelled cheque). The submission deadline may be confirmed by the issuing team.

Email: quotations@csrbox.org

Quotation must be addressed to: BharatCares

We look forward to receiving your quotation.

Regards,

ANNEXURE – I: HOSPITAL-WISE EQUIPMENT REQUIREMENTS

Civil Hospital, Manendragarh

Sr. No.	Equipment Name	Preferred Brand / Make	Specifications / Description	Warranty	Quantity
1	ICU Bed – 5 Function	Dynamic	DHF-911 Qdvance linear actuator system [04Pcs Motor, 01 Pc Control box, 01 Pc Handset] Collapsible side railing Heavy ABS Head and Foot Panel with locking 4 section top made up of oval perforated CRCA M.S. sheet Framework made up of M.S. tubes Bed mounted on 5" ABS Castors, durable and quiet Holder for Urine Bag Sitting position adjustable Finish: Pre-treated & Epoxy Powder Coated Overall approx. size: 2090L × 900W x 550-800H mm Positions obtained by Hand Remote: Hi-Low Backrest elevation Kneerest elevation Trendelenburg / Reverse Trendelenburg	1 year	10
2	Bedsheets & Pillow Covers	Abosrbant and water proof - cotton/poly cotton Indian Made	200-300 GSM	-	30
3	Blankets	Indian made	Double layered -70% wool	-	10
4	Blanket Covers	Indian made	Cotton blend	-	30
5	Basic Trolley	Yashka		1 year	1
6	Multi-Parameter Monitor	Yashka	**YID-029** Screen Size: 12.1inch TFT LCD display Resolution - 800 x600 pixels Measured Parameters - ECG,NIBP, SpO2, Rispiratory rate, Temp Alarms - Aidible and Visual Alarms Data Storage - 72hours 220V-240V, 50/60Hz 1 Year warranty	1 year	5
7	Automated External Defibrillator (AED)	Mestos	Vital Wave – Portable Biphasic Defibrillator- Event trend and storage up to 24 hours Upgradable to SpO_2 and NIBP Easy energy selection up to 270 joules Provides 50+ shocks on a fully charged battery Charges to 270 joules in < 8 seconds 7-inch color TFT display ECG acquisition through paddles and patient cable	1 year	2

8	Cold Storage Units	Not specified	Branded Voltas/Godrej or similar	1 year	1
9	Newborn resuscitation kit	Medisafe international	Specifications: Silicone Resuscitation Child (1600 ml) Guedel Airways Sizes 0,1,2, No. Silicone Face Mask Oxygen Reservoir Bagwith Valve Oxygen Tube Foot Suction with Silicone Tube	1 year	2
10	Neonatal Incubator	Neokraft-INC 1002	Modes of Operation: Servo Skin, Air, Pre-Warm & Manual Temperature setting: 19°C to 40°C, Temperature accuracy ±2% Safety: Independent high-temperature cut-off facility Canopy: Double-wall clear acrylic canopy with four port holes, baby front-loading system Inlets: Four inlets for administration of O_2, IV lines and ventilator tubings Baby tray: Tilting facility (head-up and head-down) Blower: Centrifugal blower for fresh air circulation Humidity: Humidity concentration control system with digital setting and digital display Alarms: High and low alarm facility for humidity control IRIS port: For accessing the baby without opening the canopy hood Storage: Lower drawer facility for utility purpose"	1 year	1
11	Phototherapy Unit	Dr Smith	Product Type-Phototherapy Unit LED Single Surface Lamp LED Frequency 50 Hz	1 year	2
12	Obsetric Surgical instruments: 1. Sponge Holding Forceps 2. Allis Tissue Forceps 3. Babcock's Forceps 4. Vulsellum Forceps 5. Scalpel Handles (No. 3 & No. 4) 6. Mayo Scissors (Curved/Straight) 7. Metzenbaum Scissors 8. Episiotomy Scissors	RSI	Indian Made Material- Stainless steel, rust-proof forceps with sleek, tapered tips for precise gripping. Durable, easy to clean, and perfectly balanced for delicate handling.	1 year	1

	9. Artery Forceps (Straight/Curved) 10. Kocher's Forceps (with teeth) 11.Needle Holder (Mayo-Hegar) 12. Dissecting Forceps (Toothed/Untoothed)				
13	Bedside Locker	Dynamic	DHWE-1102 Patient's Medicine Cupboard (STD) Description Overall size: 400L × 400W × 810H mm Jointless S.S. top with three sides raised edges One box with door made of M.S. sheet Four MS legs fitted with PVC stumps Finish: Pre-treated and Epoxy Powder Coated Size - 450*450*760 One Drawer One Towel rail Four 50mm Castors	1 year	6
14	Handwashing / Sanitization Stations	Dynamic	2/3 tap Made of S.S. 304, 18 SWG, with a matt finish. Surface is easy to clean. Water temperature is adjustable. Includes automatic timer control sensor. Uses a foot-operated valve. Has a thermostatic pressure-regulating mixing valve (one per station). Includes check-stop strainers. Water timer controller sensor available if required. Optional high-mount swivel gooseneck spout with foot-pedal soap dispenser (on request).	1 year	2
15	OT Table Electrical	Dynamic	DHT-3303	1 year	1
16	Surgical Instrument Set	RSI	**Material - German Steel**	1 year	1
17	Surgical Lights	Dynamic	Technical Specifications (OT Light) Intensity: 1,60,000 + 1,60,000 Lux ±10% Size of Light Field: 15 cm Colour Temperature: 3800–4800 K Colour Rendering Index (CRI): 93 RA	1 year	3

			LED Life: 50,000 hours Number of LEDs: 48 + 48 (16 Warm + 32 Cool White) Diameter of Light: 700 mm + 700 mm Brightness Control: Capacitive Touch Panel Power Supply: 220V / 50Hz AC Focusing: Adjustable Battery Backup: Optional		
18	Electrosurgical unit	Dynamic		1 year	1
19	Trauma beds	Dynamic	**DHPT 1150** **Two-section top, height adjustable by foot-operated imported hydraulic pump (from 660 mm to 910 mm)** **Gas spring assisted Trendelenburg / Reverse Trendelenburg positions** **Equipped with corner buffers, synthetic rubber covered handles, sliding X-ray cassette, accessories tray, oxygen cylinder holder** **S.S. telescopic IV rod and swing-away S.S. side rails** **Finish: Pre-treated and Epoxy Powder Coated** **Four swivel castors, 5" dia, two with total lock** **X-ray translucent removable stretcher, backrest raised on ratchet** **Overall size:**	1 year	10
20	Portable Ventilator	Resmed	**Model Astral 150** **Circuit Compatibility** Single circuit with leak Single circuit with valve Double circuit **Therapy Modes** Valve Circuit Therapy CPAP (A)CV P(A)CV P-SIMV V-SIMV PS Leak Circuit Therapy CPAP (S)T P(A)C PS with SVt iVAPS* AutoEPAP** (only in iVAPS mode)		

			Mouthpiece Ventilation (A)CV P(A)CV PS/SVt **Other Therapy Features** **Manual breath** Sigh breath (recruitment) Apnea ventilation (Volume and Pressure) Preset Programs 4 preset programs available Pressure & Volume Settings Delivered Pressure: 2–50 cmH_2O (leak circuit) 0–50 cmH_2O (valve circuit) **Delivered Tidal Volume:** Adult: 100–2500 mL Pediatric: 50–300 mL Respiratory Rate: Adult: Off, 2–50 bpm Pediatric: Off, 5–80 bpm Oxygen Compatibility Supplemental oxygen flow rate up to 30 L/min Oxygen concentration varies with leak, settings, and patient effort FiO_2 monitoring: Available (built-in) **Monitoring & Data** SpO_2 monitoring: Optional FiO_2 monitoring: Included On-screen data: 30-day summary data Real-time pressure & flow (25 Hz) **ResScan Download:** 365-day summary 7-day detailed data AirView (with RCM): 365-day summary 7-day detailed data **Alarms** Wide range of configurable high/medium/low priority alarms FiO_2 and SpO_2 alarms (when sensors connected) Adjustable disconnection alarm High-pressure alarm threshold: 80 cmH_2O **Physical Specifications** Weight: 3.2 kg Dimensions: 285 mm × 215 mm × 93 mm Internal Battery: Lithium-ion, 14.4 V, 6.6 Ah (95 Wh) Up to 8 hours backup Sound Levels Sound power level: 43 ± 3 dBA Sound pressure level: 35 ± 3 dBA		

			Compliance & Safety IEC 60601-1 Classification: Class II (double insulated) Type BF Continuous operation Suitable for use with oxygen		
21	X-Ray Machine	Allengers	Allengers-100 with APR Technical Specifications X-Ray Generator Generator Type: Compact, mobile X-ray generator Output: 100 mA kV Range: Typically 40–100 kV (as per model) mA Range: Up to 100 mA Timer Settings: mAs and exposure time adjustment Supports routine radiographic exposures X-Ray Tube Tube Type: Light-weight, mobile X-ray tube Focal Spot: According to model configuration Anode: Stationary anode Mechanical Features Mobile Trolley-based design Lightweight and easy to maneuver Tube head rotation for flexible positioning Counter-balanced arm for smooth vertical and horizontal movements Power Requirements Power Supply: Standard single-phase AC supply Stabilized input recommended	1 year	1
22	Biochemistry Analyzer	BIOGENY	Modes (9) End Point, Rate, Fixed Time, Two-Point Non-Linear, Sample Blank, Multi-Standard, Sample Blank Non-Linear, Coagulation, Absorbance Calibration Types (7) Factor, Linear, Multi-Point Non-Linear (Point to Point, Cubic Spline, Logit Log) Programmable Tests: 120 Light Source:12V, 20W Quartz Halogen Lamp Measurement Methods:Monochromatic and Bichromatic Wavelengths (Filter Wheel):7 Filters: 340, 405, 505, 546, 578, 620, 670 nm 1 Free Position Range of Absorbance-0 – 3.0 Abs Resolution:0.0001 Delay / Incubation Time-3–999 seconds Read Time:30–999 seconds Flowcell:Quartz Flowcell Flowcell Volume:32 µl Display:7-inch Colour Touch Screen Printer:In-built Thermal Printer Memory:FLASH Storage Capacity:25,000 Results Result Reports:Collated by ID, Date, and Test	1 year	1

			Quality Controls: Daily and Monthly for 3 levels with LJ Plot Aspiration Volume:200–1400 µl Flowcell Temperature:37°C (Accuracy ±0.1°C, Peltier Control) Operating Voltage:220–240 V, 60 Hz Online UPS 0.5–1.0 KVA Operating Temperature:10–35°C Operating Humidity:≤80% RH		
23	Centrifuge machine	Indian	Capacity- 4x15 ml No of tubes-4 Max speed-3500 RPM WxDxM=295x295x25	1 year	1
24	BMW Trolley	Dynamic	Model - DHF - 2316 Size - 60 Ltr (20*3) 3 Colour segregation Foot operated system for touchless opening of lid Interchangable positions	1 year	5

Civil Hospital, Raigarh

Sr. No.	Equipment Name	Preferred Brand / Make	Specifications / Description	Warranty	Quantity
1	IV Stand / Saline Stand	Dynamic	Five Pronged base fitted with 5 cms Castors S.S. adjustable I.V. Rod with double/four hook (A) S.S. Pipe with plastic base (B) M.S. Pipe with plastic base (C)S.S. Pipe with S.S. base (D) M.S. Pipe with M.S. base	1 year	25
2	Operating Stool / Chair	Dynamic	Cushioned seat and back Pneumatic height adjustable Height adjustable and swinging back rest Legs chrome plated base with foot ring Base fitted with 50 mm dia castors	1 year	20
3	Mobile Lamp / Examination Light	Dynamic	Arm- Flexible Arm with on/off switch and incorporated electronic transformer Length of Arm - 750 mm Voltage -210–230V Fitted With-LED 7 × 1W / 20 × 1W / 12 × 1W / 18 × 1W Dome- MS Powder Coated	1 year	10
4	Dressing / Instrument Trolley	Dr smith-SS Hospital Dressing Trolley-11062	Material-Stainless Steel Color-Silver Number Of Wheels-4 Wheels Country of Origin- Made in India Load Capacity - 50 Kg Feature -Height Adjustable Overall Size: 750 mm × 450 mm × 810 mm	1 year	10

5	Suction Apparatus (Mobile)	yuwell	Suction Machine 7E-A Manufacturer -YUWELL Country of Origin-India Item model number 7E-A Product Dimensions- 250 mm × 200 mm × 200 mm; 4.49 kg ASIN-B0821LVDMY	1 year	5
6	Neonatal incubator (Mobile)	Neokraft- INC 1002	**Modes of Operation:** Servo Skin, Air, Pre-Warm & Manual **Temperature setting:** 19°C to 40°C, Temperature accuracy ±2% **Safety:** Independent high-temperature cut-off facility **Canopy**: Double-wall clear acrylic canopy with four port holes, baby front-loading system **Inlets:** Four inlets for administration of O_2, IV lines and ventilator tubings **Baby tray:** Tilting facility (head-up and head-down) **Blower**: Centrifugal blower for fresh air circulation **Humidity:** Humidity concentration control system with digital setting and digital display **Alarms:** High and low alarm facility for humidity control **IRIS port:** For accessing the baby without opening the canopy hood **Storage:** Lower drawer facility for utility purpose	1 year	1
7	Incubator Hood / Darkening Cover		To be included as component of Neonatal Incubator		
8	Special Mattress		To be included as component of Neonatal Incubator		
9	Positioning Rolls / Nests		To be included as component of Neonatal Incubator		
10	Low-Pressure Suction Unit	YUWELL	7E-A Portable Phlegm Suction Unit is a newly designed oil-free lubrication pump based on extensive design experience and development trend of domestic and overseas products. It is specially designed for absorbing thick liquid such as blood, phlegm, etc. It is not only used as an aspirator in operation rooms, and emergency rooms, but also at home as a healthcare product. High negative pressure, large flux, low noise New mechanism design with suction bottle hidden inside and square negative pressure meter crust made of plastic Design reduces working noise and makes unit more fashionable and precise Adopts advanced overflow-protection design preventing pus/mucus from leaking into the machine	1 year	1

			Compact, lightweight, easy to carry — ideal for emergency & outdoor use		
11	Obsetric Surgical instruments: 1. Sponge Holding Forceps 2. Allis Tissue Forceps 3. Babcock's Forceps 4. Vulsellum Forceps 5. Scalpel Handles (No. 3 & No. 4) 6. Mayo Scissors (Curved/Straight) 7. Metzenbaum Scissors 8. Episiotomy Scissors 9. Artery Forceps (Straight/Curved) 10. Kocher’s Forceps (with teeth) 11.Needle Holder (Mayo-Hegar) 12. Dissecting Forceps (Toothed/Untoothed)	RSI	Indian Made Material- Stainless steel, rust-proof forceps with sleek, tapered tips for precise gripping.Durable, easy to clean, and perfectly balanced for delicate handling.	1 year	3
12	Patient Beds	Dynamic	**DHF-924** ABS moulded head and foot panel with locking and corner buffer Framework made of M.S. tubes Two-section top made of perforated CRCA M.S. sheet Legs fitted with PVC stumps Four locations for I.V. rod Finish: Pre-treated and epoxy powder coated Overall size: 2000 L × 900 W × 600 H mm Positions obtained by smooth crank system with foldable handle: Backrest elevation (Covers IV Poles, Mattresses, Head and Foot Board with IV Slots)	1 year	200
13	IV Stand / Saline Stand	Dynamic	**DHF-2002** S.S. telescopic I.V. rod	Not specified	include with beds

14	Mattresses & Overlays	Dynamic	**DHF-2001B** 2-section mattress	Not specified	include with beds
15	Head / Foot Board with IV slot	Dynamic	**DHF-926** S.S. head & foot panels	Not specified	include with beds
16	Bedsheets & Pillow Covers	Abosrbant and water proof - cotton/poly cotton	200-300 GSM	1 year	600
17	Blankets	Double layered -70% wool	Double layered -70% wool	1 year	600
18	Curtain / Privacy Screen	Dynamic	**DHPT - 1028** With curtain Folding arrangement included Total folds: 3 Middle section constructed in one piece Number of hinges for joining panels: 4 Frame material: CRCA pipes, 19.05 mm outer diameter, 1.22 mm thickness Base material: CRCA pipes, 25 mm outer diameter, 1.22 mm thickness Castor wheels: Plastic Hinges material: Mild Steel (MS) Hooks material: Stainless Steel (SS 202) Overall height: 1700 mm Overall width: 2360 mm Side panel width: 610 mm Middle panel width: Not applicable Curtain type: Coated fabric designed to resist bacterial adherence Curtain material: Cotton Warranty period: 1 year Test reports required by buyer: Not provided OEM ISO certifications: ISO 9001: Yes ISO 14001: Yes ISO 18001: Yes BIFMA certification: Yes	1 year	60

19	Bedside Locker	Dynamic	DHWE-1102 Patient's Medicine Cupboard (STD) Description Overall size: 400L × 400W × 810H mm Jointless S.S. top with three sides raised edges One box with door made of M.S. sheet Four MS legs fitted with PVC stumps Finish: Pre-treated and Epoxy Powder Coated	1 year	100
20	Consultation desks	Indian made-White and Brown Plywood Wooden Doctor Office Table	Primary Material-Plywood Material-Wooden Shape-Rectangular Color-White and Brown Storage-With Storage Storage Type-Drawer Storage Finish-Powder Coated Usage/Application-Doctor Office Size / Dimensions-4×2 Feet No. of Drawers-3 Drawers Table Height-4 Feet Frame Material-Mild Steel Size 4×2 Feet 4x2feet	1 year	10
21	Handwashing / Sanitization Stations	Dynamic	**DHA - 3313** Made of S.S. 304, 18 SWG, with a matt finish. Surface is easy to clean. Water temperature is adjustable. Includes automatic timer control sensor. Uses a foot-operated valve. Has a thermostatic pressure-regulating mixing valve (one per station). Includes check-stop strainers. Water timer controller sensor available if required. Optional high-mount swivel gooseneck spout with foot-pedal soap dispenser (on request). Installation is simple. Available in different bay sizes: 700 mm × 625 mm × 1360 mm (operating height 910 mm) 1400 mm × 625 mm × 1360 mm (operating height 910 mm) 2100 mm × 625 mm × 1360 mm (operating height 910 mm) Available in manual, semi-automatic, fully automatic, and automatic models with hot and cold water arrangement.	1 year	1

22	Wheelchairs	Dynamic International/Drsmith	**DWC -1174** Double Cross Bar Fixed Arm Rest PVC pad Fixed footrest, flip up plate, with leg strap 8'' solid front wheel 24'' solid rear wheel Locking breaks on rear wheel Pocket on the back	1 year	10
23	Screen Partitions	Dynamic	**DHPT - 1028** With curtain Folding arrangement included Total folds: 3 Middle section constructed in one piece Number of hinges for joining panels: 4 Frame material: CRCA pipes, 19.05 mm outer diameter, 1.22 mm thickness Base material: CRCA pipes, 25 mm outer diameter, 1.22 mm thickness Castor wheels: Plastic Hinges material: Mild Steel (MS) Hooks material: Stainless Steel (SS 202) Overall height: 1700 mm Overall width: 2360 mm Side panel width: 610 mm Middle panel width: Not applicable Curtain type: Coated fabric designed to resist bacterial adherence Curtain material: Cotton Warranty period: 1 year Test reports required by buyer: Not provided OEM ISO certifications: ISO 9001: Yes ISO 14001: Yes ISO 18001: Yes BIFMA certification: Yes	1 year	10
24	Stretchers	Dynamic	**DHPT - 1153** Frame work made of 1.25" x 16G vertical and 1" x 18G horizontal CRC tubes Trolley mounted on 15 cms dia castors, 2 with brakes Removable stretcher top made of M.S. sheet with S.S. handle Overall size: 213L x 56W x 81H cms Finish: Pre-treated and Epoxy Powder Coated Mattress Side railings M.S./S.S.	1 year	10
25	Trauma beds	Dynamic	**DHPT 1150** Two-section top, height adjustable by foot-operated imported hydraulic pump (from 660 mm to 910 mm) Gas spring assisted Trendelenburg / Reverse Trendelenburg positions	1 year	10

			Equipped with corner buffers, synthetic rubber covered handles, sliding X-ray cassette, accessories tray, oxygen cylinder holder S.S. telescopic IV rod and swing-away S.S. side rails Finish: Pre-treated and Epoxy Powder Coated Four swivel castors, 5" dia, two with total lock X-ray translucent removable stretcher, backrest raised on ratchet		
26	Mattresses & Overlays	Dynamic	standard accessory with the bed	1 year	include with beds
27	IV Stand / Saline Stand	Dynamic	standard accessory with the bed	1 year	include with beds
28	Crash Cart	Dynamic	**Model - DHF-1182** Overall approximate size: 960 L × 500 W × 1545 H mm Framework made of S.S. tubes Two lightweight polystyrene boxes, each with three drawers Upper drawer with medicine containers of different sizes Provision to hold oxygen cylinder made of M.S. Massage board for cardiac patients provided Six colour-coded hand-out small bins at eye level Total three S.S. shelves with railings to carry medical equipment Provided with S.S. I.V. Rod and corner buffers Cart mounted on 125 mm diameter castors (2 with brakes)	1 year	4
29	Water Cooler with Inbuilt RO	Blue Star	**SWCNST80120UVE**	1 year	2

Pt Deen Dayal Upadhyay Medical College, Churu

Sr. No.	Equipment Name	Preferred Brand / Make	Specifications / Description	Warranty	Quantity
1	100 kg Commercial Laundry Machine		INDUSTRIAL WASHER EXTRACTOR - HIGH CAPACITY STANDARD SOFT MOUNT - SWX 100 INDUSTRIAL TUMBLE DRYER - M.S - O3ELECTRIC - STDME 100	2.0	1
2	Pharmacy Shelving	Godrej	22 open and 22 closed	1.0	44

2. GENERAL SCOPE OF SUPPLY

• Supply of equipment and furniture as per the approved make, model and specifications.

• Transportation and safe delivery to the respective hospital locations.

• Installation, assembly and commissioning wherever applicable.

• Demonstration and basic operational training to hospital staff for medical equipment.

• Providing warranty certificates, user manuals, technical documents and relevant quality certifications.

• Providing after-sales service and support during the warranty period.

3. VENDOR QUOTATION FORMAT

Vendors are requested to provide their commercial quotation in the following format. Please provide a separate quotation line for every item and clearly identify the hospital for which it is being quoted.

Sr. No.	Hospital	Equipment Name	Make / Model Offered	Qty.	Unit Rate (■)	GST (%)	Total Amount (■)
1							
2							
3							

Note: All quantities mentioned are subject to final approval and confirmation by BharatCares. Items for which brand, specification, warranty or quantity is not specified in the source equipment sheet have been marked accordingly or left for vendor clarification. Vendors must clearly identify any items requiring clarification before submitting the final quotation.